

HS_05dec2022 on NEO500



Current Data Parameters
NAME HS_05dec2022
EXPNO 10
PROCNO 1

F2 - Acquisition Parameters
Date_ 20221207
Time 16.25 h
INSTRUM CAB AV4 500 MHZ BASIC
PROBHD Z150364_0007 (
PULPROG zg
TD 66560
SOLVENT CD3CN
NS 64
DS 0
SWH 10416.667 Hz
FIDRES 0.313001 Hz
AQ 3.1948800 sec
RG 101
DW 48.000 usec
DE 34.36 usec
TE 298.0 K
D1 10.00000000 sec
TD0 1
SFO1 500.0760004 MHz
NUC1 1H
P1 12.00 usec
PLW1 14.55099964 W

F2 - Processing parameters
SI 131072
SF 500.0730000 MHz
WDW EM
SSB 0
LB 0.10 Hz
GB 0
PC 1.00

